

Certificate of Analysis

IMMUNOSCAN CCPlus®

REF RA-96 PLUS

LOT XS 3096


2019-12-15

Reagents			LOT		Specification	Result
Ag			XS 1993	2020-09-07		
CAL	A		XS 2727	2020-04-10	OD	2.07
CAL	B		XS 2728	2020-04-10	OD	1.73
CAL	C		XS 2729	2020-04-10	OD	1.04
CAL	D		XS 2730	2020-04-10	OD	0.39
CAL	E		XS 2731	2020-04-10	OD ≥0.15	0.22
CONTROL	REF		XS 2732	2020-04-10	OD ≥0.15	0.21
CONTROL	+		XS 2725	2020-04-10	180-340 U/mL	215
					Ratio 4.2-7.6	5.1
CONTROL	-		XS 2726	2020-10-05	OD	0.01
					Ratio <0.95	0.05
DIL			XS 2721	2020-10-05	OD	0.05
CONJ			XS 2720	2020-04-15		
BUF	WASH	20X	XS 2723	2020-10-13		
SUBS		TMB	XS 2719	2020-08-31		
H ₂ SO ₄		0.5M	XS 2448	2020-08-31		

It is hereby certified that this LOT has been manufactured, tested and approved according to our quality procedures.

Date of approval: 2018-12-14



Quality Approved


Svar Life Science AB
Postal address

P.O. Box 50117
SE-202 11 Malmö
Sweden

Visiting address

Lundavägen 151
SE-212 24 Malmö
Sweden

www.svarlifescience.com